

Answers to requests for clarification RFP22-3453 - 6 May 2022

Lack of answers to specific requests, please propose a desired model in order to offer you the same or an equivalent. Thank you

Sorry we cannot give proper reference number or brand for the fairness of the process and also because some equipment references are not define yet.

NO	ITEM DESCRIPTIONS	CLARIFICATIONS	Estimated QTY per year	SPC responses
3	Pipette stands for single channel microliter pipettes	For which brand of pipette - How many storage spaces - Rotary or bench	1	compatible with item 23, holds 6 single pipettes
7	Analytical Weighing Balance. Taille du plateau > 90mm, automatic internal calibration, precision 0.1mg, range 0.2g - 300g	With or without weighing cage	1	With
8	Refrigerated benchtop centrifuge, max 15000rpm, temp range -10 to +40°C,	For which centrifugation volume, which adapters required	2	4x400mL with a swing-out rotor, 6x94mL with a fixed-angled rotor
19	Pure water system, provide type I, II, and III water	daily need in the different types of water	1	Daily use for type 1, but the same machines should also make type 2 and 3. Max 10L in a day
22	8-channel electronic pipettor , ergonomic design, highly reproducible	A specific volume - Variable or fixed - A specific brand?	1	fixed volume per pipette
23	Electronic pipettor 1 channel, lightweight and compact, comfortable handling	A specific volume - Variable or fixed - A specific brand?	1	fixed volume per pipette
27	Vortex thermo-shaker	a volume to shake - For which glassware	1	for 24 x 1.5ml tubes (it should be the default block)
32	Bottle-top dispenser/ dispenser 0.5 to 5 mL, autoclavable, ISO 8655 standards, increments 0.1mL	for which liquid: acid, base, nitric acid, sulfuric ...	1	RNA later, ethanol
54	Laboratory freeze-dryer with wide range of configurations to adapt it optimally to our specific needs.	volume freeze-dry - What are your specifics needs	1	Freeze dryer (volume: 140 diameter, 285mm height) and starter product package : 3 shelves dia. 200mm = 920cm ² , distance between shelves 85mm, additional ports for 8 flasks, round-bottom flasks and wide-neck bottles or distributor
56	Safety shower, foot and hand operation, eyewash station	To hang on the wall or a whole that rests on the floor	1	You can provide quote for both
58	Desk mount LED lab magnifier lamp	A specific magnification	2	2
62	Gravity Convection Oven, 65L, 50° to 250°C	How many shelves	1	2
65	Chainman belt + cotton thread	Length --?	1	Up to 5 km
66	Chain man cotton thread	Length	3	5 km
75	Tips 20µL, filter, steril, non pyrogenic, great recovery, in rack	For which model/brand of pipette	2	tips fit all models of the same volume
76	Tips 200µL, filter, steril, non pyrogenic, great recovery, in rack	For which model/brand of pipette	2	tips fit all models of the same volume
77	Tips 1000µL, filter, steril, non pyrogenic, great recovery, in rack	For which model/brand of pipette	2	tips fit all models of the same volume
82	Labpure S2 Purification Cartridge - SDI Feed	For which model of device	2	compatible with item 19
83	0.05 Micron ultra-microfilter (UMF)	For which model of device	2	compatible with item 19
84	Labwater Disinfection Tablets	A reference specific / conditionning	1	compatible with item 19
90	Test tube rack, PP	What volume of tube and how much space per rack	10	for 50mL falcon type tubes
92	Calibration solution pHmeter, Ph Buffer 4.0, 250mL	certified or not - In bottle or singlet - Accept a volume alternative ?	10	Certified is preferable but not essential; bottle; volume alternative is okay
93	Calibration solution pHmeter, pH buffer 7.0, 250mL	certified or not - In bottle or singlet - Accept a volume alternative ?	10	Certified is preferable but not essential; bottle; volume alternative is okay
94	Calibration solution pHmeter, pH buffer 10.0, 250mL	certified or not - In bottle or singlet - Accept a volume alternative ?	10	Certified is preferable but not essential; bottle; volume alternative is okay
95	KCL electrolyte solutions, Type L 2114, 2 mol/l KNO3 + 0.001 mol/l KCl for Ag-probes, 250 ml	certified or not - In bottle or singlet - Accept a volume alternative ?	5	Certified is preferable but not essential; bottle; volume alternative is okay
99	Erlenmeyer flask, 1000mL borosilicate	What type of collar, with or without graduation	10	With graduation, without lid sharpening
100	Erlenmeyer flask, 250mL borosilicate	What type of collar, with or without graduation	10	With graduation, without lid sharpening
101	Flask, 250ml borosilicate	With cap, what accuracy class	10	Class A
139	Plastic bags, 40x60mm	A specific colour, thickness or pictogram	100	Natural - clear, 50µm
140	Plastic bags, 100x150mm	A specific colour, thickness or pictogram	100	Natural - clear, 50µm
141	Plastic bags, 200x300mm	A specific colour, thickness or pictogram	100	Natural - clear, 50µm
142	Microtube Storage Box, assorted (Blue, Green, Orange, Purple, Yellow)	How many microtubes to store per storage	1	96 and/or 100
146	Latex Gloves S	Powdered or not	3	Not powered
147	Latex Gloves M	Powdered or not	3	Not powered
148	Latex Gloves L	Powdered or not	3	Not powered
149	Latex Gloves XL	Powdered or not	3	Not powered